

TEST REPORT

1 of 1

MODEL: DT-9881 Particle Counter

Serial Number:171201160

Report Number:201735573

Reference Instrument: FLUKE985

Instrument Series Number:1210993188

Environment Temperature:23±3℃

Humidity: 50±20%RH

Calibration Date:2017-12-13

Issue Date:2017-12-13

Calibrated By: *Li Menglong*

Issue Name: Peng Xingen

Calibrate Type	Display Value	Test Value	Result
HCHO Measure	±5%F.S	±5%F.S	Passed
CO Measure	±5%F.S	±5%F.S	Passed
Repeatability Testing	≤10%FS	<10%FS	Passed
Distribute Deviation of Particle	±30%	<30%	Passed
Particle Density Value Deviation	±30%FS	<30%FS	Passed
Air Flow	2.83L±5%	2.80L	Pass

Conclusion: Pass at Calibration Item.